



Contribution ID: 439

Type: **Poster**

## Ultralow background Kapton and Kapton-copper laminates

Signal sensor and readout cables are a fundamental component of rare event and ultrasensitive detectors. While possessing unique electrical and mechanical properties, flexible cables can be a significant contributor to the total detector background. Intrinsic contaminations of Th-232 and U-238 in commercially available flexible cables have been measured in the mBq/kg range, which can be significant compared to the levels required for next-generation rare event detectors. As a first step toward fabricating ultralow background cables, we investigated the possibility of obtaining low radioactivity (microBq/kg range) polyimides and copper-polyimide laminates for use in flexible cables. We have determined the dominant source of contamination in Kapton and demonstrated the ability to produce radiopure Kapton and Kapton-based laminates.

### Mini-abstract

Clean Kapton: a gateway to low background flexible cables?

### Experiment/Collaboration

**Primary author:** DI VACRI, Maria Laura (Pacific Northwest National Laboratory)

**Co-authors:** ARNQUIST, Isaac (Pacific Northwest National Laboratory); SALDANHA, Richard (Pacific Northwest National Laboratory)

**Presenter:** DI VACRI, Maria Laura (Pacific Northwest National Laboratory)

**Session Classification:** Poster session 3