

XENON GC/MS Analysis Neutrino Platform

R. Guida, B. Mandelli, M. Corbetta



EP-DT

Detector Technologies

STATUS

- detector > 700'000 kg **liquid Argon**
- 1kg Xenon injected in liquid Argon >> 0.00015 % of Xe = **1.5 ppm of Xe**
- Xenon bottle from Linde, purity 5.0 >> **max impurities 10 ppm**
- possible impurity in liquid Argon 1.5 ppb

- *Linde* declares a set of impurities, each lower than 1-2 ppm
- Ar, Kr should not affect the detector; CO₂, O₂, H₂O should be adsorbed by dedicated filters (on liquid phase, 700'000 kg recirculated every week)

BOTTLE IMPURITIES

- Impurities declared by Linde for Xenon 5.0

[→ Produktdatenblatt](#)



Xenon 5.0



Reinheit in %: $\geq 99,999$

Nebenbestandteile, ppm:		
N ₂		≤ 1
O ₂		≤ 0,5
H ₂ O		≤ 2
KW		≤ 0,5
H ₂		≤ 1
Ar		≤ 1
CF ₄		≤ 1
Kr		≤ 1
CO + CO ₂		≤ 1
C ₂ F ₆		≤ 1


Angaben sind als ideale Volumenanteile (= Molanteile) zu verstehen

GC/MS ANALYSIS

- Xenon is sent to Gas Chromatograph, three columns are available: PPU, OV, MolSiev
- OV1 and PPU column can be coupled with Mass Spectrometer
- Different temperatures were tried for OV1 and PPU but with similar results, best method is standard RPCMethod (95°C)
- First analysis are done with plastic pipe as connection, then moved to inox pipe to limit possible Air/O₂/N₂/H₂O adsorption

Method: C:\Soprane\Method\LAB256_Xenon

Module	<input checked="" type="checkbox"/> Module A OV1	<input checked="" type="checkbox"/> Module B PPU	<input checked="" type="checkbox"/> Module C MS5A
Inlet temp. (°C)	<input type="checkbox"/>		
Inject temp. (°C)	<input checked="" type="checkbox"/> 70.00	<input checked="" type="checkbox"/> 80.00	<input checked="" type="checkbox"/> 85.00
Column temp. (°C)	<input checked="" type="checkbox"/> 95.00	<input checked="" type="checkbox"/> 95.00	<input checked="" type="checkbox"/> 105.00
Pump (sampling time) (s)	Pump1: 60.00		
Sampling time (s)	60.00	60.00	60.00
Inject time (ms)	25.00	25.00	100.00
Backflush time (s)			
Run time (s)	240.00		
Column pressure (psi)	<input checked="" type="checkbox"/> 28.00	<input checked="" type="checkbox"/> 28.00	<input checked="" type="checkbox"/> 33.00
Detector	<input checked="" type="checkbox"/> ON	<input checked="" type="checkbox"/> ON	<input checked="" type="checkbox"/> ON




GC/MS ANALYSIS

- Chromatograms can be used to quantify elements with peak area, but a calibration is needed > rough estimation with “fraction method”
- Proportionality between area and concentration is linear up to 20%
- Retention time of a peak in the GC is the same as in the MS
> we can check peak correspondence


GC/MS ANALYSIS

- Both coupling are tested, still the most relevant is with PPU as it can separate possible impurities we are looking for > **CF₄**, **C₂F₆**
- Mass spectrometer analysis are repeated for different mass ranges, as reducing the mass range can help improving instrument sensitivity to low concentrations (expected ~ fraction of PPM)
- NIST library available in MS software for identification
> only possible if all mass lines are identified, i.e. not in small range


GC ANALYSIS : OV1 COLUMN plastic



RT 15.5s > Air
RT 16.5s > Xe
RT 18.5s > H₂O





GC ANALYSIS : OV1 COLUMN inox




MS ANALYSIS : OV₁ COUPLING

full mass range > identified Xenon, found lines 28/32 (Air) and 18(H₂O)






MS ANALYSIS : OV₁ COUPLING

mass range 10-15 > identified H₂O (RT same as GC)






GC ANALYSIS : PPU COLUMN plastic



RT 18.7s > Air
RT 22.4s > Xe
RT 108.9s > Ion69?

RT 20.27s > unidentified
RT 21.47s > unidentified
RT 21.78s > unidentified

GC ANALYSIS : PPU COLUMN inox





RT 18.7s > Air
RT 22.4s > Xe

RT 20.27s > unidentified
RT 21.47s > unidentified
RT 21.78s > unidentified

look at same RT in MS



MS ANALYSIS : PPU COUPLING

full mass range > identified Air and Xenon



MS ANALYSIS : PPU COUPLING


mass range 33-125 (out Air and Xe)
 > identified lower line of Xe and Argon



MS ANALYSIS : PPU COUPLING


mass range 70-128 (out Air, Ar and Xe, possibly in SF6 (127))

> identified lower line of Xe, Krypton



MS ANALYSIS : PPU COUPLING

specific search for CF₄/C₂F₆ > range around mass 69 (CF₄, C₂F₆) and 119 (C₂F₆)




from MS relative abundance
 $119/69 = 0.41$



>> presence of C₂F₆

MS ANALYSIS : PPU COUPLING

specific search for SF6 > range around mass 89, 108




GC ANALYSIS : MS COLUMN plastic



O2 and N2 peaks identified
not Air ratio (1:4)
> surely extra O2
> could be also extra N2
calibration needed to quantify
> 1k ppm > plastic pipe!

GC ANALYSIS : MS COLUMN inox



inox = almost no Air intake
O₂ and N₂ peaks identified
still not Air ratio (1:4)
>> extra O₂ is present

quantification done from
(very) old calibration
O₂ ~ 100 ppm

no coupling available with MS column...

CONCLUSION

- **Impurities found in the Xe 5.0 bottle:**
 - Argon, Krypton : expected, not relevant
 - H₂O : only with plastic pipe = analysis pollution
 - **O₂ : expected but in low conc. (<1ppm), quantified from GC around 100ppm**
 - **C₂F₆ : expected in low conc. (< 1ppm)**
also identified with IR spettroscopy!
- **Ongoing:**
 - proper calibration for O₂ quantification with GC
 - quantification of C₂F₆ with MS, not straightforward + calibration needed
 - analysis of previously used Xe bottle > no problems for detector prototype