

## MAX IV 3 GeV NEG coated storage ring status

*Thursday, 18 April 2024 11:25 (25 minutes)*

The vacuum system of the 3 GeV storage ring at MAX IV laboratory is based on a fully NEG (Non-evaporable Getter) coated, extruded copper pipe (22 mm inside diameter) that also serves as a distributed absorber intercepting synchrotron light. The ring has been in operation since the end of 2015 and performs very well. This presentation is summarizing the design and performance of the 3 GeV storage ring.

### Summary

**Primary author:** Mr GRABSKI, Marek (MAX IV Laboratory)

**Co-authors:** Mrs AL-DMOUR, Eshraq (MAX IV Laboratory); Mrs MARTINEZ CARBONERES, Ana (MAX IV Laboratory); Mr SCOLARI, Simone Maria (MAX IV Laboratory)

**Presenter:** Mr GRABSKI, Marek (MAX IV Laboratory)

**Session Classification:** session 5