

 BERKELEY LAB	Lawrence Berkeley National Laboratory <h1 style="margin: 0;">Nonconformance Report (NCR)</h1> MQXFA-NCR-0276	LBNL Document # SU-1013-7818	Rev A	Page 1 of 3
--	---	--	-----------------	----------------

Date: 10/7/21 QA Doc #: **MQXFA-NCR-0276**

Originator:	K. Ray	Company:	BNL
Phone:	510-486-4024	PO Number:	
E-mail:	klray@lbl.gov	Other:	

Part Name: Reference Coil Quantity Impacted: 2

Part Number: 27K950 Serial No. (if applicable): 220, 221

Additional Notes:

ITEM NO.	NONCONFORMING CHARACTERISTIC	DISPOSITION	REINSPECTION	
1	Some pins were sticking into the slot at the coil pole.	<input type="checkbox"/> ACCEPT AS-IS <input checked="" type="checkbox"/> REWORK <input type="checkbox"/> REJECT Trim the pins	<input checked="" type="checkbox"/> YES <input type="checkbox"/> NO	
2		<input type="checkbox"/> ACCEPT AS-IS <input type="checkbox"/> REWORK <input type="checkbox"/> REJECT	<input type="checkbox"/> YES <input type="checkbox"/> NO	
3		<input type="checkbox"/> ACCEPT AS-IS <input type="checkbox"/> REWORK <input type="checkbox"/> REJECT	<input type="checkbox"/> YES <input type="checkbox"/> NO	

Use continuation sheet if necessary. Attach red-lines, photos as necessary

APPROVAL:	CONCURRENCE:	CONCURRENCE:	CONCURRENCE:
Technical Lead Date	QA Rep. Date	Stakeholder/CAM Date	PM (High Risk NCRs Only))

Indicate all follow-on or related processes: NONE CAR ECN HOLD

Is it a 'Suspect/Counterfeit' Item? ☐ YES* ☒ NO

*Report to OCA (Office of Contractor Assurance) and follow PUB-3111 to dispose material

Date: 10/7/21 QA Doc #: **MQXFA-NCR-0276**



	<div>Lawrence Berkeley National Laboratory</div> <div>Nonconformance Report (NCR)</div> <div>MQXFA-NCR-0276</div>	<div>LBNL Document #</div> <div>SU-1013-7818</div>	<div>Rev</div> <div>A</div>	<div>Page</div> <div>3 of 3</div>
---	---	---	------------------------------------	-----------------------------------

Date: 10/7/21 QA Doc #: **MQXFA-NCR-0276**

