

# Micro Pattern Gaseous Detectors (MPGDs) for Nuclear Physics

F. Barbosa<sup>a</sup>, M. Cortesi<sup>c</sup>, S. Dalla Torre<sup>1</sup>, Y. Furletova<sup>a</sup>, K. Gnanvo (Lead Organiser)<sup>a,\*</sup>, M. Hohlmann<sup>f</sup>, D. Neyret<sup>d</sup>, M. Posik (Lead Organiser)<sup>b</sup>

<sup>a</sup>*Thomas Jefferson National Accelerator Facility, Newport News, VA 23606, USA*

<sup>b</sup>*Temple University, Philadelphia, PA 23606, USA*

<sup>c</sup>*Facility for Rare Isotope Beams, Michigan State University, East Lansing, MI 48824, USA*

<sup>d</sup>*CEA Saclay IRFU, Université Paris-Saclay, France*

<sup>e</sup>*Florida Institute of Technology, Melbourne, FL 32901, USA*

<sup>f</sup>*INFN Trieste*

---

## Contents

<b>1</b>	<b>Executive Summary</b>	<b>4</b>
1.1	MPGD Technologies for Nuclear Physics at Jefferson Lab . . . . .	4
1.2	MPGD Technologies for Low Energy Nuclear Physics at FRIB . . . . .	4
1.3	Advanced Micro Pattern Gas Detectors for Tracking at the Electron Ion Collider . . . . .	4
1.4	MPGD Technologies for Particle Identification in Nuclear Physics Experiments . . . . .	4
1.5	Electronics, DAQ, and Readout Systems for MPGD Technologies . . . . .	5
1.6	Need for Dedicated Nuclear Physics MPGD Development Facility . . . . .	5
<b>2</b>	<b>Advanced Micro Pattern Gas Detectors for Tracking at the Electron Ion Collider</b>	<b>6</b>
2.1	The Electron Ion Collider . . . . .	6
2.1.1	Accelerator and Physics Overview . . . . .	6
2.1.2	MPGD Tracking Requirements at the EIC . . . . .	6
2.2	Recent and Current R&D Efforts . . . . .	6
2.2.1	Accommodating a Small Material Budget . . . . .	6
2.2.2	Readout Structures . . . . .	7
2.3	Future MPGD R&D related to EIC . . . . .	8
2.4	Need for a Nuclear Physics MPGD Development Facility . . . . .	8

---

\*Corresponding author

Email address: kagnanvo@jlab.org (K. Gnanvo (Lead Organiser))

<b>3</b>	<b>MPGD technologies for Low Energies Nuclear Physics</b>	<b>10</b>
3.1	Introduction . . . . .	10
3.2	Requirements of low-energy nuclear physics experiments at FRIB . . . . .	10
3.2.1	Rate and multiplicity . . . . .	10
3.2.2	Gas Gain and dynamic range . . . . .	11
3.2.3	Size and complexity and versatility . . . . .	11
3.3	Present MPGDs-based detectors at FRIB . . . . .	11
3.3.1	Drift chamber for tracing at the focal plane of high-rigidity spectrometers . . . . .	11
3.3.2	Time-Projection-Chambers operated in Active-Target mode (AC-TPC) . . . . .	11
3.3.3	The Gaseous Detector with Germanium Tagging (GADGET) . . . . .	12
3.3.4	Path Forward and new opportunities: . . . . .	12
3.3.5	Broader Impact: . . . . .	13
3.3.6	Common facility for production and development of new fabrication technologies . . . . .	13
<b>4</b>	<b>Development of Large MPGDs for High-Rate Experiment at Jefferson Lab</b>	<b>14</b>
4.1	Introduction . . . . .	14
4.2	MPGD technologies at JLab . . . . .	14
4.2.1	Overview of Current MPGDs in Experiment at JLab . . . . .	14
4.2.2	MPGDs Needs for Future Experiments at JLab: Common aspects & differences with HEP . . . . .	14
4.3	Challenges and R&D efforts for future experiments at JLab . . . . .	15
4.3.1	Tracking in High Rate Environment . . . . .	15
4.3.2	Development of low channel count & high performance readout structures . . . . .	16
4.4	Need for MPGDs R&D facility in the US for the Nuclear Physics community . . . . .	17
<b>5</b>	<b>MPGD Technologies for Particle Identification in Nuclear Physics Experiments</b>	<b>18</b>
5.1	The role of MPGD-based Photon Detectors in RICH Technologies . . . . .	18
5.2	MPGD-based Transition Radiation Detector . . . . .	20
5.2.1	Physics motivation . . . . .	20
5.2.2	Current experience/experiments . . . . .	20
5.2.3	R&D for future experiments ( prototype,readout, radiator,gas) . . . . .	21
5.3	PICOSEC: MPGD for TOF at the EIC . . . . .	23
<b>6</b>	<b>Electronics, DAQ &amp; readout system for MPGD technologies</b>	<b>24</b>
6.1	Streaming readout at the EIC . . . . .	24

6.2	Streaming readout for MPGD detectors . . . . .	25
6.2.1	Online MPGD data reduction and filtering . . . . .	25
6.2.2	ML techniques for MPGD data processing . . . . .	25
6.3	MPGD front-end electronics adapted to triggerless mode . . . . .	25
6.3.1	Presently existing front-end ASICs . . . . .	25
6.3.2	Project for a new chip in 65 nm technology . . . . .	26
6.3.3	Needs for specific MPGD applications . . . . .	27

## 1. Executive Summary

Many major nuclear physics (NP) experiments across the United States are developing MPGD detectors to be used for tracking and PID purposes. The unique environment of NP experiments are driving the R&D requirements to develop detectors and their electronics which are ideally suited for NP.

### 1.1. MPGD Technologies for Nuclear Physics at Jefferson Lab

Future spectrometer for NP experiments at the Thomas Jefferson National Accelerator Facility (Jefferson Lab) require large area  $O(m)$ , low mass ( $X/X_0 \leq 1\%$ ), excellent spatial  $O(100\mu m)$ , excellent timing  $O(10\text{ ns})$ , high rate  $O(1\text{MHz}/\text{cm}^2)$  tracking detectors technologies for operation in high background rate and high radiation environment. Only MPGD technologies such as Gas Electron Multipliers (GEMs), Micro Mesh Gaseous (Micromegas) or Resistive Micro Well ( $\mu\text{RWELL}$ ) detectors are able to satisfy the challenges of high performances for large acceptance at reasonably low cost. Critical R&D for the next decades will focus on new ideas to develop ultra-low mass and large area and radiation tolerant MPGD trackers while ensuring the high rate capabilities. Performance of novel material (Chromium GEMs, Aluminum based readout strips) will be investigated as well as original concepts for anode readout such as capacitive and resistive and zigzag readouts to develop low channel count and high spatial resolutions MPGD detectors.

### 1.2. MPGD Technologies for Low Energy Nuclear Physics at FRIB

The Facility for Rare Isotope Beams (FRIB) at Michigan State University (MSU) will become the world's most advanced facility for the production of rare isotope beams (RIBs). With the delivery of beams starting in Spring 2022, FRIB will be capable of producing a majority (around 80%) of the isotopes predicted to exist including more than 3,000 new isotopes, opening exciting perspectives for exploring the uncharted regions of the nuclear landscape. Its scientific impact will span a better understanding of open quantum systems at the limits of stability through investigations of the structure and reactions of atomic nuclei and their roles in nuclear astrophysics, low-energy tests of fundamental symmetries, and practical applications that benefit humanity. MPGD technologies play an important role for the success of the science program at FRIB. Applications of MPGD technologies include low-pressure tracking and particle-identification (PID) at the focal planes of magnetic spectrometers, Active-Target Time-Projection-Chambers (TPCs), and TPCs for the detection of exotic decay modes with stopped RIBs. The unprecedented discovery potential of FRIB can be achieved by implementing state-of-the-art experimental equipment and overcoming challenges of present devices by taking the following measures: improving spatial and energy resolutions, optimizing pure-gas operation for active target mode, improving reliability and radiation hardness at a lower cost, reducing ion-back flow to minimize secondary effects and increase counting rate capability, and integrating electronic readout to reach high channel density, fast data processing and storage.

### 1.3. Advanced Micro Pattern Gas Detectors for Tracking at the Electron Ion Collider

Successfully completing the physics program of the Electron-Ion Collider (EIC), to be built at Brookhaven National Laboratory, requires its tracking system to have low mass ( $X/X_0 \lesssim 1\%$ ), large area  $O(1m)$ , and excellent spatial resolution  $O(100\mu m)$ . MPGDs, such as GEMs, Micromegas, and  $\mu\text{RWELL}$  can meet these requirements. Furthermore, the EIC is expected to have relatively low rates, below  $100\text{ kHz}/\text{cm}^2$ , which is well within the operating range of current MPGDs. Current R&D is focused on reducing large-area detector material budgets and reducing the number of channels needed to be read out while maintaining excellent spatial resolution. The EIC can benefit from future R&D which looks to further reduce the material and service budget of the detector, pushing the spatial resolution down towards  $O(20\mu m)$ , and implementing particle identification capabilities into MPGD detectors.

### 1.4. MPGD Technologies for Particle Identification in Nuclear Physics Experiments

Particle identification (PID) plays a very important role for HEP and NP physics. The next generation of high intensity accelerators and high demand for precision measurements from the physics will require the development of high granularity detectors, such as those based on MPGD technologies. Combining a high precision tracker with PID capable technology could prove valuable for future experiments. A high precision MPGD tracker combined with a

transition radiation (TR) option for particle identification could provide important information necessary for electron identification and hadron suppression. A radiator, installed in front of a MPGD entrance window provides an efficient yield of TR photons. Coupling MPGD technology with a Cerenkov radiator equipped with a photocathode, a time-of-flight (TOF) detector can be created, which can be used to PID. MPGD technology lends itself well to covering large areas, allowing for hermetic TOF detector.

### *1.5. Electronics, DAQ, and Readout Systems for MPGD Technologies*

The EIC will implement a full streaming readout architecture. This trigger-less implementation will consist of front-end circuitry and processors to enable data collection, processing and analysis: front-end ASICs will be designed to meet wide bandwidth sub-detector and system requirements; front-end processors will include FPGAs to provide data aggregation and enable flexible algorithms to reduce data volume while maintaining wide system bandwidth; system clock distribution with timing precision of the order of 1 ps; link exchange modules and servers for data processing; and data transport via extensive use of optical fibers. The development of ML/AI algorithms will play a critical role in enabling a full detector bandwidth of 100 Tbps and to deliver data output rates of 100 Gbps. Readout of MPGD detectors requires specific front-end ASICs able to amplify and digitize the detector signals with performance requirements depending of its constraints and application. The ASICs should be also compatible with the high-speed streaming readout DAQ systems that are considered for the future experiments at EIC and elsewhere. Present chips, like SAMPA or VMM, partially satisfy these requirements and can be used for specific applications. Nevertheless, an initiative is launched to develop a new versatile ASIC covering most of the constraints of the different MPGD applications in HEP experiments. The new SALSA chip is meant to be equally adapted to the streaming and triggered readout paradigms and is being designed in a more modern 65 nm technology. In addition, the SALSA chip holds a promise to be the base of a new family of integrated circuits covering several specific applications requiring on-demand adaptation of some of its functionalities.

### *1.6. Need for Dedicated Nuclear Physics MPGD Development Facility*

As highlighted throughout this white paper, the requirements of detectors for NP experiments differ from those in HEP, which leads to different detector R&D paths and priorities. A dedicated US-based facility for the development, testing and dissemination of MPGD technologies is strongly needed within the US NP MPGD community. Implementing a MPGD facility in the US similar to the Gaseous Detector Development (GDD) facility at CERN or the SiDet facility at FNAL, which provided silicon detector development for the HEP community in the US, for to the development of MPGD technologies would be beneficial for not only the nuclear physics community, but also for the US particle physics community in general, such as HEP or astrophysics groups. Such MPGD facility could be based at one of DOE's National Laboratories, such as Jefferson Lab or Brookhaven National Laboratory.

## 2. Advanced Micro Pattern Gas Detectors for Tracking at the Electron Ion Collider

### 2.1. The Electron Ion Collider

#### 2.1.1. Accelerator and Physics Overview

The Electron-Ion Collider (EIC) is a new facility being built at Brookhaven National Laboratory (BNL) in the U.S [1, 2]. The EIC is quite unique compared to other colliders, such as the LHC and RHIC. It will collide polarized electrons and polarized protons/ions at high luminosity ( $10^{33} - 10^{34} \text{ cm}^{-2} \text{ s}^{-1}$ ) and cover a wide range in center of mass energies ( $\sqrt{s_{ep}} = 28 - 140 \text{ GeV}$ ). The collisions of beam species (ep/eA) lead to backgrounds such as hadron beam backgrounds (e.g. beam gas events) and synchrotron radiation. The beam energies will be asymmetric resulting in boosted kinematics, with higher activity occurring at larger pseudorapidity ( $\eta > 1$ , the hadron going direction). The bunch spacing will be  $\sim 9 \text{ ns}$  and the beams will collide with a 25 mrad crossing angle. Another distinguishing feature of the EIC relative to other colliders is that it will have low particle multiplicity ( $\lesssim 10$  tracks/event), with an interaction rate of 500 kHz and insignificant pileup. Finally, the radiation environment will be much lower compared to the LHC (factor of  $\sim 100$ ). Taking advantage of the high beam polarizations, large luminosity, and wide kinematic coverage, the EIC will make high precision measurements over a broad range of physics topics [2] via deep inelastic scattering (DIS), semi-inclusive DIS (SIDIS), and exclusive processes. To achieve this, an EIC detector must have a large pseudorapidity coverage ( $|\eta| < 4$ ), and far forward subdetectors capable of going beyond  $\eta = 4$ ,  $4\pi$  coverage, a high-precision and low-mass tracking system, and excellent particle identification (PID) performance to separate  $e$ ,  $\pi$ ,  $K$ , and  $p$  at the track level.

#### 2.1.2. MPGD Tracking Requirements at the EIC

The unique environment of the EIC sets the tracking detector requirements, which differ significantly from other collider experiments. To provide the momentum resolution required by the physics program, the material budget available for tracking detectors is rather low, specifically below 1%  $X/X_0$  per layer. This is needed to minimize multiple scattering in order to preserve electron and photon measurements. Finally, large-area tracking detectors can be used to ensure that the full kinematic acceptance is covered. Micropattern Gas Detectors (MPGDs), e.g. GEM, Micromegas, and  $\mu$ rwell, are well suited for an EIC tracking system as they can be designed to meet these EIC tracking requirements. Furthermore, the EIC is expected to have particle rates well below  $100 \text{ kHz/cm}^2$ , which current MPGDs are capable of handling. The following subsections will summarize the current R&D being carried out to produce large-area and low-mass MPGD trackers (Sec.2.2.1), high resolution readout structures suited for the EIC (Sec. 2.2.2), and future R&D (Sec. 2.3).

### 2.2. Recent and Current R&D Efforts

In January 2011, Brookhaven National Laboratory (BNL), in association with Jefferson Lab and the DOE Office of Nuclear Physics established a generic detector R&D program [3] to address the requirements for measurements at the EIC. From this program, the eRD3 and eRD6 projects were formed to focus on EIC compatible MPGD detectors, whose efforts have contributed greatly towards realizing high-resolution, large-area, and low-mass tracking detectors [4, 5, 6, 7, 8, 9, 10, 11, 12, 13, 14]. Current experiments such as Jefferson Lab's Super Bigbite Spectrometer (SBS) and CLAS12 experiments have already benefitted from this R&D work. With the EIC moving towards becoming a reality after clearing CD0 and CD1 DOE milestones, a more targeted EIC R&D program has been established with the eRD108 project focusing on further developing MPGD based detectors for the EIC experiment [15]. These R&D activities have been focusing on two main aspects: reducing the material budget of large-area detectors and developing low-channel-count, but high-resolution readout structures.

#### 2.2.1. Accommodating a Small Material Budget

To meet the momentum resolution and kinematic coverage requirements needed to successfully complete the EIC physics program, it is critical to reduce as much as possible the material of the tracking system, not only in the active

area of the detectors, but also in its support structures. Passive structures are typically located within the tracking volume and can contribute to multiple scattering and acceptance holes. Building large-area detector modules or tiles reduces the amount of support material located in the tracking volume. However, building large-area detector modules while satisfying a low material budget is a non-trivial task since when detectors become larger they typically require more support. MPGD detectors, such as triple-GEMs are typically supported with G10/FR4 support material. A current R&D task of eRD108 is investigating new materials which have lower radiation length and offer stronger tensile support.

One way to reduce the active area of an MPGD tracker is to keep the standard amplification foils, such as GEM or  $\mu$ RWELL foils, but design the cathode and readout elements as Cu/Kapton based foils. Following this procedure, eRD6 was able to successfully build triple-GEM detectors with a length of roughly  $1m$  and radiation length of  $\sim 0.4\% X/X_0$  and operate and test then in beam. Applying this approach to building a  $\mu$ RWELL detector will further reduce the material budget since the material of the amplification element of a  $\mu$ RWELL is less than that of a triple-GEM, and one can use the same foil-based cathode and readout elements as the triple-GEM detector.

A significant contribution to the material budget of MPGD detectors is the use of Cu in the electrodes, which has a relatively large radiation length. The charge sensitive elements of the readout are usually made of Cu strips (or pads). Replacing these Cu strips with Al strips will yield a decrease in the detector material. This is being investigated within eRD108, specifically applied to Micromegas. Additionally, eRD108 is also investigating reducing the traditional woven Inox mesh of the Micromegas, which is used to create the electron amplification, with thin ( $\lesssim 40\mu m$ ) layers of Al foils. If successful, this would reduce further the Micromegas material budget.

### 2.2.2. Readout Structures

*Channel Reduction.* The EIC will require high spatial hit resolution and large-area detectors, which traditionally is achieved through finely pitched charge-sensitive strips or pads, which for large-area detectors would require lots of channels to be read out. This not only makes building, operating, and analyzing the detector more cumbersome due to its many channels, but also significantly increases its cost. There are several R&D efforts ongoing, which began with eRD6 and are now being further developed by eRD108, to lower the required channel count while keeping excellent resolution. The goal for this R&D is to obtain  $100\mu m$  spatial resolution for large detectors ( $\sim 5,000 cm^2$ ) using a readout strip pitch of  $1 mm$ . One way to achieve this is to exploit the charge-sharing behavior of particular readout strip structures used to read out the detector. Already demonstrated is using a 1D zigzag strip structure [16, 4, 7], to achieve a better resolution than suggested by the wide pitch of the strips via the sensitive charge sharing among the strips. Being able to have a coarser strip pitch, but achieve the resolution of straight strips with a finer pitch reduces the number of detector readout channels. R&D activities [16, 4, 7] testing 1D zigzag strip readouts on small MPGD prototypes ( $\sim 100 cm^2$ ) have already demonstrated spatial resolutions around  $50\text{--}100 \mu m$  are achievable with up to  $2 mm$  strip pitches. Scaling up the detector size and implementing a 2D version of the zigzag strip structure are being investigated.

Developing flexible capacitive sharing readout structures is another way that the channel count of a detector could be reduced while maintaining excellent spatial resolution. These readout structures consist of a vertical stack of alternating layers of charge-sensitive strips (or pads) and Kapton. As one moves away from the amplification region, the pitch of the strips becomes larger with each layer, with only the bottom layer (widest pitch) being read out. The various layers are capacitively coupled, so a charge in the top layer (closest to the amplification layer) will induce a charge in the layer below it, which in turn induces charge in the next layer below, and so on. The resolution of the readout is then set by the narrow strip pitch of the top layer, but the number of channels needed to be read out is set by the number of strips in the wider strip pitch of the bottom layer. There is a diamond-like carbon (DLC) resistive layer on top of the structure which serves to sink the charges from the amplification and contributes to spreading out the initial charge to aid the charge sharing among strips. Initial tests using a small scale triple-GEM prototype ( $\sim 100 cm^2$ ) equipped with a 2D capacitive sharing pad readout which used  $0.52 mm \times 0.52 mm$  pads with a pitch of  $6.125 mm \times 6.125 mm$  structure as the top layer, and readout the bottom layer which consisted of  $9.99 mm \times 9.99 mm$  pads with a pitch of  $10 mm \times 10 mm$ . A spatial resolution of around  $250 \mu m$  [17]. Ongoing R&D efforts within eRD108 are pursuing capacitive sharing strip readouts and equipping large area ( $\sim 5,000 cm^2$ ) detectors with them.

*Resolving Track Multiplicity Ambiguities.* The EIC can also benefit from readout structure R&D related to minimizing ‘ghost hits’ or track multiplicity ambiguities. While the EIC is not expected to have a large track multiplicity ( $\lesssim 10$  tracks/event), being able to resolve such ambiguities will aid in the track reconstruction efficiency and performance. One solution would be to make use of a three-coordinate readout structure to provide a correlation between three coordinates. These three coordinate structures are well suited for implementing into a capacitive sharing readout structure [17]. Another option to investigate would be the use of fast electronics, for example ASICs with a faster sampling rate, to use the timing of the hits to remove or separate ambiguous hits.

### 2.3. Future MPGD R&D related to EIC

The tracking performance of MPGD detectors can be improved by further reducing their material to minimize multiple scattering and dead areas that are in the tracking acceptance. The spatial resolution of the readout is limited by the spatial resolution inherent to the MPGD detector itself, which is driven mainly by fluctuations on the position of the ionization cluster in the drift gap. Investigating the size of the drift gap and constructing detectors with several narrower drift gaps could lead to better spatial resolution.

Although excellent momentum resolution via tracking at the EIC is critical to completing its physics program, so is the PID ability of an EIC detector. Implementing tracking detectors that are also capable of performing PID would be beneficial to the EIC. Detectors such as MPGD-based transition radiation detectors (TRDs) and time-of-flight (TOF) detectors would make excellent additions to the EIC providing both hit points and PID capabilities (see Sec. ?? and IF5 White Paper 1). These types of detectors would rely on the development of fast electronics (e.g. sampling rate  $\sim 100\text{MHz}$ ) at reasonable cost per channel and MPGDs with excellent timing resolutions ( $\lesssim 1\text{ns}$ ).

### 2.4. Need for a Nuclear Physics MPGD Development Facility

The MPGD tracking requirements for the EIC require large-area, low mass, and high precision space point resolution. Furthermore, the EIC would greatly benefit from PID capable MPGD detectors, which could provide large-area and low-mass detectors capable of providing high precision hit points with PID information. Implementing a dedicated facility in the US for the development, testing, and dissemination of MPGD technologies will strongly benefit the development of MPGDs needed for the EIC. Such a facility could be modeled on the Gaseous Detector Development (GDD) facility at CERN or the SiDet facility at FNAL, which is based in the US and develops silicon detectors for HEP. Although the GDD facility at CERN serves the global MPGD community, due to their geographical proximity and strong affiliation to the CERN-based RD51 collaboration for MPGD technologies, MPGD groups involved in HEP experiments at CERN, European universities and research institutions have benefited more. A similar facility based in the US would also serve as a center for the US MPGD community to coalesce around, strengthening what is already a growing community of detector experts.

## References

- [1] J. Adam et al. Electron-ion collider at brookhaven national laboratory: Conceptual Design Report. [https://www.bnl.gov/ec/files/EIC\\_CDR\\_Final.pdf](https://www.bnl.gov/ec/files/EIC_CDR_Final.pdf), 2021.
- [2] R. Abdul Khalek et al. Science requirements and detector concepts for the electron-ion collider: EIC Yellow Report, arXiv:2103.05419, 2021.
- [3] Generic detector r&d for an electron ion collider. [https://wiki.bnl.gov/conferences/index.php/EIC\\_R/25Df](https://wiki.bnl.gov/conferences/index.php/EIC_R/25Df), 2011-2021.
- [4] B. Azmoun, S. Aune, K. Dehmelt, A. Deshpande, W. Fan, P. Garg, T. K. Hemmick, M. Kebbiri, A. Kiselev, I. Mandjavidze, H. Pereira-Da-Costa, C. E. Perez-Lara, M. L. Purschke, M. Revolle, M. Vandenbroucke, and C. Woody. Design studies of high-resolution readout planes using zigzags with gem detectors. *IEEE Transactions on Nuclear Science*, 67(8):1869–1876, 2020.
- [5] M. Vandenbroucke, H. Alves, S. Aune, B. Azmoun, Klaus Dehmelt, A. Deshpande, Prakhar Garg, T. Hemmick, M. Kebbiri, A. Kiselev, Irakli Mandjavidze, H. Costa, C. Lara, M. Purschke, and C. Woody. A study of “zigzag” strip readout for micromegas detectors. pages 1–4, 11 2018.
- [6] B. Azmoun, S. Aune, Klaus Dehmelt, Abhay Deshpande, W. Fan, Prakhar Garg, T. Hemmick, M. Kebbiri, A. Kiselev, Irakli Mandjavidze, H. Pereira-Da-Costa, C. Perez-Lara, M. Purschke, M. Revolle, Maxence Vandenbroucke, and C. Woody. Design studies of high resolution readout planes using zigzags with gem detectors. *IEEE Transactions on Nuclear Science*, PP:1–1, 06 2020.



- [7] C. Perez-Lara, S. Aune, B. Azmoun, K. Dehmelt, A. Deshpande, W. Fan, P. Garg, T. K. Hemmick, M. Kebbiri, A. Kiselev, I. Mandjavidze, M. L. Purschke, M. Revolle, M. Vandenbroucke, and C. Woody. A comparative study of straight-strip and zigzag-interleaved anode patterns for mpd readouts. *IEEE Transactions on Nuclear Science*, 69(1):50–55, 2022.
- [8] Marcus Hohlmann, Matthew Bomberger, Stefano Colafranceschi, Francisco Jimenez, Mehdi Rahmani, and Aiwu Zhang. Low-mass GEM detector with radial zigzag readout strips for forward tracking at the EIC. In *2017 IEEE Nuclear Science Symposium and Medical Imaging Conference (NSS/MIC 2017) Atlanta, Georgia, USA, October 21-28, 2017*, 2017.
- [9] Aiwu Zhang, Marcus Hohlmann, Babak Azmoun, Martin L. Purschke, and Craig Woody. A GEM readout with radial zigzag strips and linear charge-sharing response. *Nucl. Instrum. Meth.*, A887:184–192, 2018.
- [10] Aiwu Zhang and Marcus Hohlmann. Accuracy of the geometric-mean method for determining spatial resolutions of tracking detectors in the presence of multiple Coulomb scattering. *JINST*, 11(06):P06012, 2016.
- [11] Aiwu Zhang, Vallary Bhopatkar, Marcus Hohlmann, Xinzhan Bai, Kondo Gnanvo, Nilanga K. Liyanage, Matt Posik, and Bernd Surrow. R&D on GEM detectors for forward tracking at a future Electron-Ion Collider. In *Proceedings, 2015 IEEE Nuclear Science Symposium and Medical Imaging Conference (NSS/MIC 2015): San Diego, California, United States*, page 7581965, 2016.
- [12] Aiwu Zhang, Vallary Bhopatkar, Eric Hansen, Marcus Hohlmann, Shreeya Khanal, Michael Phipps, Elizabeth Starling, Jessie Twigger, and Kimberly Walton. Performance of a Large-area GEM Detector Read Out with Wide Radial Zigzag Strips. *Nucl. Instrum. Meth.*, A811:30–41, 2016.
- [13] Kondo Gnanvo, Nilanga Liyanage, Vladimir Nelyubin, Kiadtisak Saenboonruang, Seth Sacher, and Bogdan Wojtsekhowski. Large Size GEM for Super Bigbite Spectrometer (SBS) Polarimeter for Hall A 12 GeV program at JLab. *Nucl. Instrum. Meth.*, A782:77–86, 2015.
- [14] Kondo Gnanvo, Xinzhan Bai, Chao Gu, Nilanga Liyanage, Vladimir Nelyubin, and Yuxiang Zhao. Performance in test beam of a large-area and light-weight GEM detector with 2D stereo-angle (UV) strip readout. *Nucl. Instrum. Meth.*, A808:83–92, 2016.
- [15] eRD108. Eic mpd needs and r&d strategies. [https://indico.cern.ch/event/1063927/contributions/4560105/subcontributions/356566/attachments/2344971/3998579/eRD108\\_CERN-EIC-RD.pdf](https://indico.cern.ch/event/1063927/contributions/4560105/subcontributions/356566/attachments/2344971/3998579/eRD108_CERN-EIC-RD.pdf), 2021. CERN Experimental Physics R&D Day.
- [16] Aiwu Zhang, Marcus Hohlmann, Babak Azmoun, Martin L. Purschke, and Craig Woody. A gem readout with radial zigzag strips and linear charge-sharing response. *Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment*, 887:184–192, 2018.
- [17] K. Gnanvo. New results on capacitive sharing large pad readout in test beam. [https://indico.cern.ch/event/989298/contributions/4217765/attachments/2190650/3702411/KG\\_RD51\\_MiniWeek20210216.pdf](https://indico.cern.ch/event/989298/contributions/4217765/attachments/2190650/3702411/KG_RD51_MiniWeek20210216.pdf), 2021. RD51 Mini-week.

### 3. MPGD technologies for Low Energies Nuclear Physics

#### 3.1. Introduction

The Facility for Rare Isotope Beams (FRIB) at Michigan State University (MSU) is at the forefront of experimental low-energy nuclear physics. FRIB will be the world's most advanced facility for the production of rare isotope beams (RIBs), capable of making a majority (around 80%) of the isotopes predicted to exist available for experiments; this opens exciting perspectives for exploring the uncharted regions in the nuclear landscape. Its scientific impact will span from the discovery of more than 3,000 new isotopes to constraining nuclear astrophysics processes, improving our understanding of open quantum systems at the limits of stability, and allow researchers to deepen the understanding of the atomic nuclei and their role in the Universe. The unprecedented discovery potential experiment with RIBs at FRIB can be realized by implementing state-of-the-art experimental equipment that can study these isotopes at the highest rates produced.

Gaseous avalanche readout technologies, and in particular Micro-Pattern Gaseous Detectors (MPGDs), play an important role for the success of the science program at FRIB. A few examples of application of MPGDs to experimental RIBs physics include low-pressure drift chambers for tracking and particle-identification (PID) at the focal plane of high rigidity spectrometers (refs), position-sensitive readout for Time-Projection-Chambers (TPCs) operated in active-target mode with re-accelerated radioactive beams (refs), and TPCs for the detection of exotic decay modes with stopped radioactive beams (refs).

Challenges for the future MPGD technologies applied to the field of low-energy nuclear physics with RIBs include improved spatial resolution and segmentation; better reliability and radiation hardness while minimizing power and cost; low ion-back flow for minimized secondary effects and an increase counting rate capability; integrate electronic readout to reach high channel density, fast data processing and fast data storing.

#### 3.2. Requirements of low-energy nuclear physics experiments at FRIB

Originally the MPGDs were developed in the framework of high-energy physics (HEP), with the aim of overcome limitation of wire-based readouts (refs). As soon as MPGDs technologies matured and evolved in a full plethora of diversified architectures, their applications expanded to other fields of basic research and applied science, including experimental nuclear physics, astrophysics, neutrino physics, material science, neutron detection, homeland security, medical imaging and industry (refs).

MPGDs offer large flexibility and allow one to tailor the geometry and performance to satisfy specific requirements, playing a crucial role in the development of instrumentation for the next generation of particle colliders. However, the requirements of HEP experiments are fundamentally different from the ones that characterize the domain of low-energy nuclear physics with RIBs. The implementation of MPGDs outside the HEP from which they were conceived, requires often a compromise between detector optimization, performance and development effort (namely cost). Main difference between HEP experiments and RIBs physics in the low-energy domain include:

##### 3.2.1. Rate and multiplicity

In typical HEP experiments, the complexity of nucleon-nucleon collisions at relativistic energies and the occurrence of possible new phenomena is expected to modify the global characteristics of the collisions, so that charged-particle multiplicity is one of the important global parameters used to extract the interaction-dynamics. This requires high-precision measurements of detectors with a continuous sampling of recorded events, where up to thousands tracks of secondary particles are generated from each interactions. On the other hand, low-energy RIBs physics exploit a relatively simple and well understood reactions where only a few reaction products are involved in the process, including resonant elastic and inelastic scattering of protons and  $\alpha$ -particles, one / two nucleon transfer, Coulomb excitation, fusion-evaporation reactions, and exotic radioactive decay relevant to the nuclear astrophysics science, such as beta-delay charged particle emission. In addition, because of the low intensity of RIBs is many orders of

magnitude lower than that of stable beams, together with extremely small cross subsections, the typical luminosity of these experiments is relative small. As a consequence, the efficiency of the experimental setup and the triggering strategy for the electronics become a determining factor.

### 3.2.2. *Gas Gain and dynamic range*

Central tracking detectors (i.e. TPC) for the charged particle momentum measurement, electromagnetic/hadron calorimeters for energy measurement, muon tracking systems, Ring-Imaging Cerenkov (RICH) detectors for PID, and other HEP collider detectors are either based on the recording signals produced by minimum ionizing particles (MIPs), which have extremely low specific ionization density, or by the recording of photons-generated single photo-electrons from liquid/solid converters (i.e. photocathode or liquid scintillators). As a consequence, the gas avalanche readout requires relative high gas-gain to generated a measurable signals, while the detector is operated in standard gas mixtures at atmosphere pressure, more often in highly ionizing background. To achieved detection efficiency in this conditions, a suitable wide dynamic range or the suppression of the background is necessary. Typical RIBs experiments are based on the detection of heavy ions that a large amount of primary ionization electrons realized in the sensitive volume of the detector, so that moderate gas gain is generally sufficient to achieved the required sensitivity for the detection of the reactions products. However, the reactions involved particles characterized by very different mass, charged, and emitted within a wide range of energies, so that a wide dynamic range of the detector became a crucial requirements for extracting the kinematic properties of the reaction under study.

### 3.2.3. *Size and complexity and versatility*

HEP experiments are design to run for decades under more or less same operational conditions. Sizes and cost. Feasible RIBs experiments can accumulate enough statistics for a significant measurement within a reasonable beam time, typically from some days to a week. Generally multi-tasks gaseous tracking systems are employed for a diversified portfolio of beams, types of reactions, energy range that ranges from stopped beams to a few hundred of MeV/u, and operational conditions (form a few tens of Torr up to 1 atm) - large versatility is a crucial intrinsic property of these systems. Because of the low-energy domain and the short particle range, the detectors require smaller volumes compared to the ones used in HEP domain, with a much lower complexity even when couple to ancillary systems. Moreover, for all gaseous detector the choice of the counting gas that filled the vessel is a delicate matter, generally selected depending on the desired mode of operation, working conditions and type of measurement. Most chambers run in a mixture of noble gas (e.g. Ar) and a smaller fraction of a complex, polyatomic molecule, generally hydrocarbon (CH<sub>4</sub>, CF<sub>4</sub>, CO<sub>2</sub>). The first component allows multiplication at voltage bias and its chemical stable; the second one is added to absorbs photons emitted by excited atoms in the avalanche when they return to the ground state, and suppresses secondary emission allowing high gas gains before discharge. However, AT-TPC the gas filling is counting gas as well as the reaction target, purity and low pressure. The modality and properties of the low-energy nuclear experiments with RIBs require a new class of MPGD architecture, innovative MPGD-based readouts, and deidcated front-end electronics, specifically developed to satisfy the requirements of these experiments.

## 3.3. *Present MPGDs-based detectors at FRIB*

### 3.3.1. *Drift chamber for tracing at the focal plane of high-rigidity spectrometers*

### 3.3.2. *Time-Projection-Chambers operated in Active-Target ode (AC-TPC)*

The past decade has seen a rapid increase of the number of active targets implemented in time projection chambers (AT-TPC) used in low energy nuclear physics experiments with RIBs. They provide full solid-angle detection coverage,, excellent energy and angular resolutions and separation of charged particle tracks. This results in a high luminosity ideal for experiments with weak beam intensities, down to a few hundred particles per second, therefore extending the reach of nuclear studies to more than a thousand exotic and rare isotopes that can only be produced at such low intensities. To response and satisfy the demanding challenges of low-energy AT-TPC experiments with

RIBs in terms of position-sensitive readout performance, new MPGD structures are studied and developed – multi-layer THGEM configuration with mesh imbedded as inner electrodes. In addition, we are currently focusing on the implementation of a new production technology, the additive manufacturing technology for large-scale fine gas avalanche structure, as well as the implementation of new substrate materials for hybrid MPGD configuration. The scope is to reduce discharge probability for operation of the AT-TPC readout in pure elemental gas at low pressure, at a reduced ion back-flow for high rate applications. Examples and recent progress: development of the Multi-layer Thick Gaseous Electron multiplier (M-THGEM) for the position-sensitive readout of TPC operated in active-target mode (AT-TPC). The innovative M-THGEM is capable of providing stable, high-gain operation in pure elemental gas (H<sub>2</sub>, D<sub>2</sub>, He, Ar, etc.).

### 3.3.3. *The Gaseous Detector with Germanium Tagging (GADGET)*

### 3.3.4. *Path Forward and new opportunities:*

New architecture for reducing ion back-flow at high rate: A fundamental limit of gaseous detectors to both applications of Time Projection Chambers (TPCs)[ref] and Gaseous Photomultiplier (GPM) [ref] in high-rate experiments is the accumulation of slowly-drifting, avalanche-induced ions in the active gas volume [ref]. In a TPC, an intense ion back-flow leads to a build-up of positive ions in the drift region that causes space charge effects. This compromises the homogeneity of the electric field locally, causing a degradation of the localisation capability, and of the overall performance of the position-sensitive readout. In a GPM, the avalanche-induced ions that drift back to the photocathode (PC) cause severe limitations to the gas gain stability of the detector. Furthermore, an intense stream of positive ions that impact on the sensitive PC causes a substantial modification of the lattice defects pattern in the PC layer, resulting in a reduction of the electron escape length and a variation of the electron affinity. This eventually results in a substantial loss of quantum efficiency. The gas avalanche readout is the major source of positive ions that stream back to the cathode. The amount of positive charges created during the avalanche depends on several factors, including the intensity and energy of charged particles that impinge on the detector, the type and properties of the filling gas, and the ion back-flow suppression capability of the gas amplification stage. In most traditional gaseous detector configurations, including multi-wire proportional chambers, parallel-plate counters, and resistive plate chambers, almost all avalanche ions flow back to the photocathode (GPM) or to the collection region preceding the multiplier (TPC). The ion back flow (IBF) can be reduced by many orders of magnitude by incorporating a pulsed ion-gate electrode that takes advantage of the natural delay in the ions arrival to block them by switching the drift field polarity. However, this comes at the expense of a considerable dead time which limits the detector rate down to several tens of Hz [4]. Advances in photolithography and micro-processing techniques in the chip industry during the past two decades triggered a major transition in the field of gas detectors from wire structures to Micro-Pattern Gas Detector (MPGD) concepts, revolutionizing cell-size limitations for many gas detector applications. In cascaded micro-pattern detectors, the backflow of the positive ions is dramatically reduced, as a large fraction of the ions are collected in the intermediate elements [5]. For instance, in micromegas detectors a large fraction of the avalanche ions can be stopped at the micro-mesh due to the large avalanche/drift field ratio that acts as a filter [6]. In other detector configurations consisting of a cascade of two/three hole-type multipliers, ion back streaming can be reduced to a 1% level by adjusting the transfer fields between successive elements [7]. The search for new MPGD structures for a better ion backflow suppression is one of the most pursued R&D trends in MPGDs, and specific research activities are currently ongoing in several laboratories and research centers worldwide. Success in this field will pave the way to photon detectors with chemically-unstable photo-converters, e.g. Bialkali photocathodes with visible-light sensitivity, as well as excellent point resolution in large-volume TPCs operated at much higher luminosity. The impact of such devices will not only be tremendous for the nuclear/high energy physics community but also in applied radiation detection such as radiation monitoring and imaging. Development of new gaseous photomultiplier (GPM) based on hybrid MPGD-based technology. Goal: development of low-cost, high-granularity, optical readout for large-area, fast-neutron system used for invariant-mass measurements. Further development of MPGD structures as readout for tracking system capable of low-ion feedback, high-counting rate, excellent localization capability and large dynamic range. Goal: response and satisfy the demanding challenges of experiments with rare isotope beams in terms of position-sensitive readout performance.

#### *3.3.5. Broader Impact:*

Strong collaboration with HEP groups of Snowmass will be crucial for a fast advance in the field, and it will benefit a larger scientific community. It provides new relevant scientific and technological knowledge on MPGDs and their implementation, leading to a potential larger application portfolio beyond nuclear physics.

#### *3.3.6. Common facility for production and development of new fabrication technologies*

## 4. Development of Large MPGDs for High-Rate Experiment at Jefferson Lab

### 4.1. Introduction

The Thomas Jefferson National Accelerator Facility (TJNAF), also known as Jefferson Lab or JLab is a world-leading US national laboratory at the forefront of the study of the fundamental nature of nuclear matter. The facility is home to the Continuous Electron Beam Accelerator Facility (CEBAF), a world-class machine that deliver high intensity longitudinally polarized electron beams as a probe for investigating nucleon structure and nuclear physics. JLab completed the 12 GeV upgrade of the CEBAF in 2017 which allows outstanding study of nucleon structure in terms of form factors (FFs), transverse momentum distributions (TMDs), generalized parton distributions (GPDs) and structure functions in the valence quark region. CEBAF will be capable of delivering longitudinally polarized electron beams at an intensity of up to 85  $\mu\text{A}$ , which corresponds in some experiments to luminosities as high as  $10^{39}$  electron/s-nucleon/cm<sup>2</sup>. CEBAF provide the electron beam to four experimental halls, labelled Hall A, Hall B, Hall C, and Hall D. Each hall has a highly specialized spectrometers to record the products of collisions between the electron beam to a fixed target to study the interaction of the quarks and gluons that make up protons and neutrons of the nucleus.

### 4.2. MPGD technologies at JLab

#### 4.2.1. Overview of Current MPGDs in Experiment at JLab

The 12 GeV upgrade of the Continuous Electron Beam Accelerator Facility (CEBAF) at JLab has seen the substitution of drift chambers used previously for tracking in many spectrometers of the experimental halls with large area MPGD tracking detectors in order to cope with high particle rate from the increased CEBAF electron beam luminosity and to provide spatial resolution better than 100  $\mu\text{m}$  to carry out the JLab current and future ambitious nuclear physics program. The large acceptance spectrometer CLAS12 in the experimental Hall B has adopted cylindrical Micromegas detector (MVT) as part of its vertex detector in combination with silicon layers (SVT). Super Bigbite Spectrometer (SBS) experimental program in hall A has opted for large area (200 cm  $\times$  60 cm) triple-GEM detectors as part of the tracking stations of the Bigbite and SBS in the electron and hadron arms experimental setup respectively. GEM technology offers the capability required to cope with the exceptional high rate of a few hundreds kHz/cm<sup>2</sup> demanded by the experiments in hall A. Large GEM trackers were also recently used in Hall B in the Proton Radius (PRad) Experiment as well as as cylindrical readout layers in a radial TPC configuration in the BoNUS12 experiment for recoil proton tracking. Future experiments at JLab are expanding the use of MPGD technologies for tracking. MOLLER, SoLID spectrometers in Hall A heavily rely on large area and low mass GEM trackers, The TDIS collaboration is developing a novel concept of modular Time Projection Chamber (mTPC) with GEM readout for recoil proton detection in the vicinity of the target where very high background particle rate is expected. More recently, Hall B has started investigating the possibility to replace the CLAS12 forward drift chambers with the  $\mu\text{RWELL}$  technology and PRad-II experiment (an upgrade of the previous PRad) is also looking at the new MPGD technology for its tracker layer. In Hall D, GEM-based transition radiation detectors (GEM-TRD) is been developed as additional PID option to the DIRC and the electron calorimeter to improve the  $e/\pi$  separation in the forward direction of the GLUEX detector. The challenges posed by the development of all these MPGD technologies for future applications at JLab are explored in more detail in the subsection 4.3 below.

#### 4.2.2. MPGDs Needs for Future Experiments at JLab: Common aspects & differences with HEP

MPGD technologies are widely used in high energy physics collider experiments like in the 4 LHC detectors ATLAS, CMS, LHCb typically as muons chambers or as readout planes for TPC, whereas in NP experiments such as at JLab, MPGDs are mainly used as main tracking detectors in fixed target spectrometers. The requirements imposed on the detector are therefore drastically different in term of detector design, choice of material to minimise detector thickness and ensure mechanical and radiation robustness of the detector to provide stable high performance over the lifetime of the experiment in harsh radiation environment. In both HEP and NP cases, large area, high rate capabilities, radiation hardness and excellent space point resolution are common challenges to need to be solved. The key difference is the

need for low mass (or low material budget) MPGD trackers, with radiation length typically less than 1%  $X/X_0$  per tracking layer of critical importance not only to minimize both background from mostly low energy photon conversion but also to limit the impact of multiple coulomb scattering for precise momentum resolution measurement. This very stringent low mass requirement is not as critical for HEP applications as muon chambers (reduced photon background environment, muon are not sensitivity to material thickness) or as TPC anode readout planes, the anode plane is outside of the sensitive volume of the TPC. The implication in term of the design and operation of the MPGD detectors are not to be underestimated.

One example to illustrate the importance for dedicated MPGD development for NP are the development of NS2 technique for mechanical assembly of triple-GEM championed by the CMS collaboration for the CMS muon chambers upgrade, or the large Micromegas chambers for ATLAS New Small Wheel (NSW) would never be compatible for tracking in spectrometers in JLab experimental halls. The development of these two detectors designs required more than 10 years of sustained R&D program, but can only benefit application that do not require low mass detector to the level we need to achieve at JLab. It is up to the nuclear physics community to commit to the R&D effort in order to fully optimize MPGD technologies to the specific requirements imposed by the experiments.

### 4.3. Challenges and R&D efforts for future experiments at JLab

#### 4.3.1. Tracking in High Rate Environment

**Development of large area & low mass MPGD structures.** One critical consideration to achieve the momentum resolution requirements and minimize the background production to successfully complete JLab experimental nuclear physics program is to design large area tracking system with the lowest material thickness possible. This is a requirement on the detector design not only in the active area of the detector defining its acceptance but also involving the mechanical support structure at the peripheral of the detectors modules. In the spectrometers of JLab fixed target experiments, we typically design large MPGD detector modules in order to maximize the active-to-dead area ratio in the overall acceptance of the spectrometer. Designing large MPGD detectors, however requires strong support structure which in turn result in bulky and heavier structures. As an illustration, a meter-size triple-GEM detector typically has 15 mm thick and 30 mm wide G10/FR4 frames at its peripheral to keep the GEM foils stretched and flat. This is in addition to 3 mm thick honeycomb sandwiched between thin G10 skins on which the GEM and strip readout foils are stacked. Though this triple-GEM design with a radiation length of  $\sim 1\%$   $X/X_0$  in the active area, known as the COMPASS GEM was considered as a low mass detector a decade ago, the requirements in term of low mass detector for future experiments at JLab such as SoLID or CLAS12 spectrometers are more ambitious in term of low mass detector, aiming at a reduction of the radiation length by a factor 2 ( $\sim 0.4\%$   $X/X_0$ ) per tracking layer and the width of the support frames by a factor 4 with respect to the current state of the art. One way to achieve these requirement for triple-GEM detector, would be the "Only-foils" approach in the design and construction of these detectors where the honeycomb support structure in the detector active area is removed, resulting in only GEM, cathode and strip readout foils in the active area. A first "Only-foils" triple-GEM prototype with a length of roughly 1m and radiation length of  $\sim 0.45\%$   $X/X_0$  was successfully built and operated in beam to demonstrate as part of the EIC detector R&D effort (eRD6). However the main lesson learned from this first prototype is that significant effort are needed to develop new type of low mass and very strong material to support the tension of all these stretched GEM foils and insure a stable operation of the detectors. We will be investigating replacing fiberglass frames by carbon fiber structure or very narrow ceramic-based frames. A significant contribution to the material budget of MPGD detectors is the use of Cu in the electrodes, for GEM,  $\mu$ RWELL. At the level of the  $\sim 0.4\%$   $X/X_0$  radiation length under discussion, a few tens of microns of Cu layers becomes a dominant contribution of the material of the detector in the active area. Similarly, the metallic mesh of a micromegas which is based on Inox material is the major contributor to the total material thickness. The charge sensitive elements (strips or pads) of the MPGD readout structure are also usually made of Cu. Replacing the Cu in some parts of the detectors by lighter metal such as aluminum (Al) or very thin layer of chromium (Cr) will significantly reduce the detector material. Several approaches are ongoing to carry out these R&D effort with the investigation of 2D Al-strip readout development with micromegas detector, the development of Cr-capacitive-sharing readout structures for GEM,  $\mu$ RWELL.

**Development of high-rate resistive MPGDs.** The operation of MPGD structures with high gain and in spark-free or spark-protected condition was enabled in recent years with the development of resistive MPGDs such as resistive micromegas or  $\mu$ RWELL detectors. Both structures used a thin resistive layer as part of the amplification structure to quench the energy of a spark discharge and reduce the spark rate by order of magnitudes. This feature subsequently results in a dramatic reduction of the damages to the amplification structure itself, the front end electronics caused by sparks as well as reducing the long dead time to recover from spark usually observed in standard MPGDs during operation. However the evacuation to the ground of the MPGD amplification charges through the resistive layer severely reduce the rate capability of resistive MPGD structures. The charge evacuation timing characteristics depends on the surface resistivity of the layer as well as size of the detector and quickly drop a few order of magnitude for large size MPGD compared to standard non resistive MPGDs. As an example a 1-meter size triple-GEM detector is capable to easily operate in a rate environment of  $\sim 1 \text{ MHz} / \text{cm}^2$ , while  $\mu$ RWELL or resistive micromegas structures of the similar size with a typical surface resistivity of  $\sim 10 - 20 \text{ M}\Omega / \square$ , has a limitation of the rate of the order  $\sim \text{kHz} / \text{cm}^2$ . A sustained R&D effort in the future to develop large area resistive MPGD structures with high rate capabilities ( $\sim 1 \text{ MHz} / \text{cm}^2$ ) is therefore critically important for tracking in the high rate environments of JLab experimental halls. Several R&D efforts to reach high rates capability with  $\mu$ RWELL detectors by reducing the path for the amplification charges to the ground are been investigated by different groups all across the world. Novel high rate resistive MPGD structures should however preserve the spark resistant capabilities of such detectors and long term stability by a proper choice of the material and design. The development of large area resistive MPGD structures that combines high rate with spark resistant and radiation hardness capabilities will be extremely beneficial for the next generation of nuclear physics experiments at JLab but required a robust and sustain R&D effort specially within by the US nuclear physics community.

#### 4.3.2. Development of low channel count & high performance readout structures

**Low channel count readout structures:.** MPGD detectors typically require very finely pitched strips or pads readout plane, i.e. a large number of electronic channels to be read out in order to achieve high space point resolution capabilities because the small size of the charge avalanche cloud produced by the micro structure devices used by these detectors for signal amplification. This results not only in a significant cost of readout electronics for large area MPGD tracking detectors, but also the main driver of the event data size to be processed and transferred from the DAQ system to discs or other electronic storage during the experiment and for the offline analysis. in addition, the constraints in term of cooling, services and connection of large number of front-end electronics to the detector during the design, construction and operation should not be underestimated. There are several ongoing R&D efforts significantly reduce the required channel count by a factor 5, i.e. readout strip pitch typically of 1 to 2 mm for large ( $\sim 1 \text{ m}^2$ ) MPGD detectors with the goal to keep spatial resolution below than  $100 \mu\text{m}$ . The three main ideas currently under investigation are:

- **Zigzag readout:** Uses interleaves of chevron-style / zigzag shape strips to facilitate charge sharing between wide pitch adjacent strips and therefore achieve high spatial resolution performances with low channel count. The concept have been proven for 1D zigzag readout on small MPGDs prototypes with the 3 main amplification structures (GEMs, Micromegas and  $\mu$ RWELLS). The major challenge for this technology is to replicate the performances on 2D readout and to scale it to large area readout planes.
- **capacitive-sharing readout:** An alternative approach using capacitance coupling between stack of pad layers to transfer charges to large pad or strip readout. This approach has the combined advantage of being easily adaptable to all types of MPGD amplification structures (GEMs, Micromegas and  $\mu$ RWELLS) and readout patterns. Preliminary tests on small size GEM and  $\mu$ RWELL prototypes demonstrated excellent spatial resolution performances reaching  $200 \mu\text{m}$  for  $1 \text{ cm} \times 1 \text{ cm}$  pad readouts and  $60 \mu\text{m}$  for 1 mm (X-Y) strip readouts. The next step in the development of capacitive-sharing readout structure is to achieve the same level of performances for large area readout and more importantly for U-V style strip readout structures. Another critical area that would require a sustained R&D effort in years to come would be the development of low mass & low noise capacitive-sharing readout structures using thin Al-pad or Cr-pads layers for the charge transfer to meet the requirements for large area, & low mass MPGD trackers at JLab.



- **resistive readout:** The surface resistivity of the resistive layers in Micromegas or  $\mu$ RWELL detectors could also be finely tune to allow a lateral spread of charge cloud from the amplification structure to be shared between neighboring strips or pads with large pitch and to use center of gravity to reconstruct the particle position with high accuracy. The main challenge for this approach is that characteristics of the resistive layer could be tuned only for a given MPGD amplification and readout structure. In addition, a very good control of the thickness uniformity of the resistivity layer is critical for large detector to maintain uniform performance across the area

**Development of three-coordinate MPGD readout structures.** Operating large MPGD with strip readout structures in high rate environment such as at JLab will undoubtedly results in the need to address pile-up and multiple hits ambiguity (“ghost hits”) issues. The problem is amplified with the development of low channel count readout structures with that require large strip pitches as described in the previous section 4.3.2. An elegant solution to address the multiple hits ambiguity as well as improving tracking efficiency in high background rate environment is the development of three-coordinate (X-Y-U) strips readout structure to use the good time, position and charge correlation of the 3 set of strips as a way to reconstruct 2D hit position. Capacitive-sharing structures are by design, well suited for three-coordinate X-Y-U strip readout structures because of the versatility of the concept. A GEM prototype with X-Y-U strip readout will be developed to demonstrate the concept and subsequent R&D will be performed to optimize the characteristics of such readout structure to ensure high spatial resolution performances for all 3 set of strips coordinates for large area devices. If successful, the combination of high-rate capability, low channel count and 3 coordinates measurement on a single readout structure for large area detectors will represent a turning point for the development of large MPGD devices for high rate tracking in high background environment of experiment in JLab experimental halls.

#### 4.4. Need for MPGDs R&D facility in the US for the Nuclear Physics community

As highlighted throughout this white paper, the MPGD requirements for future experiments at Jefferson Lab experiments require dedicated and specific R&D program to address the challenges of developing simultaneously large area, low mass and high performance and radiation resistant MPGD technologies for tracking and PID. Implementing a dedicated facility for the development, testing and dissemination of MPGD technologies will strongly benefit the development of MPGDs for Jefferson lab needs. Such US MPGD development facility could be modeled on the Gaseous Detector Development (GDD) facility at CERN or the SiDet facility at FNAL for silicon detector development for HEP in the US. Though the GDD facility at CERN is available for any MPGD groups of the particle physics community world wide, it has so far mostly benefit MPGD groups involved of the HEP experiments at CERN and in Europe universities and research institution mainly because of the geographic proximity and the strong affiliation to the CERN-based RD51 collaboration for MPGD technologies. Jefferson Lab has in the past, vastly benefit from the knowledge and expertise available within the RD51 collaboration to develop the MPGD trackers for Super BigBite program in Hall A or BoNUS and CLAS12 in Hall B, however with the ever increasing interest using these technologies in future experiment at the lab, a US based MPGD development and testing facility coalescing the US MPGD groups to address the challenges will be critical for the success of the MPGD detectors development at Jefferson Lab. The facility could be hosted at one DOE National Laboratory such as Jefferson Lab or BNL but made available for all MPGD community including NP, HEP, astrophysics as well as for medical and industrial application.

## 5. MPGD Technologies for Particle Identification in Nuclear Physics Experiments

### 5.1. The role of MPGD-based Photon Detectors in RICH Technologies

Gaseous Photon Detectors (PD) have played/are playing a major role in establishing and operating Ring Imaging CHerenkov (RICH) counters, thanks to their specific characteristics, some of them unique: they represent the most cost effective solution for what concerns the coverage of large detector areas, and they offer the minimum material budget, a feature relevant when the photon detectors have to sit in the experiment acceptance. Moreover, the gaseous PDs can operate in presence of magnetic field.

The successful operation of gaseous PDs imposes to overcome two major challenges.

- **Selection of the photoconverter** - The photoconverting vapours, initially used, require either extended conversion volumes (TMAE), that results in parallax errors and wide ranges of electron drift time, or very far UltraViolet (UV) detection domain (TEA). Feedback photons from the multiplication process can generate spurious hits wherever in the converting volume. They have been progressively abandoned. Among the standard solid state photoconverters commonly used in vacuum-based detectors, only CsI can be reliably used in gaseous atmosphere thanks to its relatively high work function: it can tolerate some bombardment by the ions generated in the multiplication process, where the maximum integrated bombardment before observing Quantum Efficiency (QE) degradation is of the order of  $1 \text{ mC cm}^{-2}$  [1].
- **Photoelectron extraction** - In gas atmosphere, the extracted photoelectrons can be elastically back scattered by the gas molecules and be reabsorbed in the photoconverter. Effective photoelectron extraction requires specific gas atmospheres and high electric field in front of the photoconverters [2]. MultiWire Proportional Chambers (MWPC) equipped with CsI photocathodes [3] have been successfully operated, for instance in HADES, COMPASS and ALICE RICHes, even if at low gain in order to limit the ion bombardment and the photon feedback. In these detectors, where the signal is due to the ion motion, low gain results in slow operation.

MPGD technologies offer natural answers to ion back flow and photon feedback suppression and much faster operation, as tested by successful applications:

- the PHENIX HBD with triple GEM PDs [4];
- the COMPASS RICH upgrade with Hybrid (THGEMS and MICROMEGAS) PDs [5];
- the windowless RICH prototype and test beam with quintuple GEM PDs [6];
- the TPC-Cherenkov (TPCC) tracker prototype with quadruple GEM PDs [7].

In multiple layer GEM PDs, where the top layer is coated with a CsI film and acts as photocathodes, the photon feedback is stopped by the limited optical transparency of the stack of GEMs, while the ion backflow is reduced because part of the ions are trapped in the intermediate detector layers. In a hybrid detector including two THGEMS and a MICROMEGAS, the first THGEM is the photocathode substrate and the feedback photon are stopped as in the GEM detector. The intrinsic ion blocking characteristics of the MICROMEGAS makes possible photon feedback rates at a few percent level. These detectors can be used in focusing and proximity focusing RICHes.

The major element of interest for future applications of MPGD-based PDs is in developing the concept of compact RICH for PID of high momentum particles, that would empower the application at colliders with hermetic coverage detectors and, therefore, is a must at the EIC. In fact, RICHes for PID at high momenta require gaseous radiators, as only small-value refractive index give access to PID at high momenta. The radiator must be long to ensure the required Cherenkov photon yields. Higher photon rates can be obtained in the far UV domain, around 120-140 nm. Access to this wavelength range can be obtained by a windowless RICH citewindowless, where the radiator gas is also the PD gas. This poses specific requirements to MPGD-based PDs. FluoroCarbons (FC) are mainly used as radiator

gasses thanks to their high density, that ensures good Cherenkov photon rates, and their low chromaticity, that make possible fine resolution, a need for PID at very high momenta. The FC Global Warming Potential (GWP) is extremely high and, therefore, their use is subject to increasing restrictions, also effecting procurement possibilities. A proposed alternative is by pressurized (at a few bar) noble gasses able to mimic FC in terms of density and chromaticity. MPGDs operation in FC atmosphere has been proven [6], while their ability to operate in high pressure noble gasses has to be established, in spite of some positive hints from literature[8]. Another need for the compact RICH concept is the fine pixelization, required to preserve the fine resolution with shorter lever arm imposed by the compactness requirements.

The possibility to identify novel solid-state photoconverters providing higher QE and adequate for operation in gaseous PDs has to be pursued: it is beneficial for the compact RICH concept and, more in general, for all the applications of gaseous PDs. Hydrogenated nanodiamond powders have been proposed [9] and initial studies are ongoing [10], while further investigation is needed, with dedicated attention to novel C-materials.

In conclusion, MPGD-based PDs are an option for further developments in the Cherenkov imaging techniques and, in particular, for the needed concept of compact RICHes, essential at the EIC and that, more in general, can open the way to a wider use of RICHes in collider environments.

## 5.2. MPGD-based Transition Radiation Detector

### 5.2.1. Physics motivation

The electron identification plays a very important role for a physics emerging at the Electron-Ion Collider. The following processes are regarded as essential for EIC physics program and could be accessed with a help of improved electron identification. Events with electrons in the final state are important signatures for DIS physics at EIC. Secondary electrons could be emitted from leptonic and semi-leptonic decays of hadrons. It tight together with an efficiency for indentifications of such processes as  $J/\psi$  production (branching ratio to  $e^+e^-$  pair is the order of 6%), D-mesons production ( with its  $\text{Br}(D^+ \rightarrow e + X) \sim 16\%$ ), as well as B-mesons production ( lepton  $\text{Br}(B^\pm \rightarrow e + \nu + X_c) \sim 10\%$ ) [? ]. Electron identification plays an important role for many other physics topics, such as spectroscopy, beyond the standard model physics, etc. Isolated electrons are not easy to identify at the EIC because of the relatively large QCD background from hadrons.

The next generation of high intensity accelerator and high demand on precision measurements from the physics have to result on the development of high granularity detectors. A high granularity tracker combined with a transition radiation option for particle identification could provide additional information necessary for electron identification or hadron suppression.

### 5.2.2. Current experience/experiments

Transition radiation (TR) is produced by charged particles when they cross the boundary between two media with different dielectric constants [? ]. The probability to emit one photon per boundary crossing is of order  $\alpha \sim 1/137$ . To increase the transition radiation yield, multi-layer dielectric radiators are used, typically several hundred mylar foils, polyethylene foam, or fibers (fleece) [? ]. The energies of transition radiation photons emitted by relativistic particles are in the X-ray region with a detectable energy range of 3-50 keV [? ]. These photons are extremely forward peaked (within an angle of  $1/\gamma$ ). The total transition radiation energy emitted is proportional to the  $\gamma$ -factor of the charged particle.

Typically, in particle physics, transition radiation detectors (TRDs) are used for electron identification and for electron/hadron separation. Transition radiation detectors were widely used in the high energy experiments, such as ATLAS, ALICE, ZEUS, HERMES. They were based on the multi-wire chambers or straw tubes.

The initial concept and first design of GEM-based Transition Radiation detector was proposed for the EIC detector R&D program [? ]. This program is supported by the DOE Office of Nuclear Physics.

It combines a Gas Electron Multiplier (GEM) tracker with the TRD functionality. A standard GEM tracker [? ] with high granularity (400  $\mu\text{m}$  strip pitch) capable of providing high resolution tracking was converted into a transition radiation detector and tracker (GEM-TRD/T). This was achieved by making several modifications to the standard GEM tracker. First, since heavy gases are required for efficient absorption of X-rays, the operational gas mixture has been changed from an Argon based mixture to a Xenon based mixture. Secondly, the drift region also needed to be increased from  $\sim 3$  mm to 21 mm in order to detect more energetic TR photons. Then to produce the TR photons, a TR radiator was installed in front of the GEM entrance window. Finally, the standard GEM readout (based on the APV25 [? ]) was replaced with one based on the relatively faster, JLAB developed, flash ADC (FADC) [? ].

The GEM-TRD/T concept is shown in Fig. 1.

During the generic EIC R&D program a small  $10 \times 10\text{cm}$  prototype, as shown in the Fig ?? has been build and tested [? ].

The first beam test measurements using the GEM-TRD/T prototype have been performed at Jefferson Lab (CEBAF, Hall-D) using 3-6 GeV electrons.

First results show that an  $e/\pi$  rejection factor of 9 can be achieved with a single GEM-TRD/T module and can be increased up to a factor 16 by using a thicker radiator (ca. 25 cm) with 98% of electron efficiency.




Figure 1: GEM-TRD/T operation principle




Figure 2: Schematic of GEM-TRD/T prototype

### 5.2.3. R&D for future experiments ( prototype, readout, radiator, gas)

We would like to continue this effort and are happy to invite experimentalists to develop MPGD-types of TRDs for the future experiments as EIC and beyond. For the successful operation of MPGD-based TRDs the following challenges need to be addressed:

- **Large scale prototype** During the generic EIC R&D activity our group used a small-size ( $10 \times 10 \text{ cm}^2$ ) to validate a property of GEM-based TRD concept. Once the EIC project moves forward (to a large-scale production stage) one needs to perform tests on the large-size modules in order to be able to work out possible issues: like noise, gain-uniformity, drift-time issues, HV stability, etc. This would allow to work-out the design issues of gas/field cages to make sure that they are suitable for TR applications.
- **Field/gas-cage and HV optimization** In order to keep the electric field uniform a special field cage needs to be developed. This includes the mechanical design and construction of the field-/gas-cage to minimize a Xe-filled gas gap between the radiator and the drift cathode.

The GEM TRD will need 2 HV lines: one for the GEM amplification stage and the second to set a uniform drift field. To work in a high occupancy environment, the drift time needs to be minimized, requiring fields of  $\sim 2\text{-}3 \text{ kV/cm}$ . For a 2 cm drift distance the HV should be at the level of 4-5 kV. Depending on the chosen grounding scheme, the total voltage including GEM stage, could be up to 8-9 kV. Optimization of HV for large drift distances for a large scale prototypes needs to be performed.

- **Anode readout PCB layers** Different anode readout PCB layers could be evaluated, which includes capacitive-sharing pad readout, large size 2D strip readout of zigzag readout options. One needs to demonstrate that the concept works equally well with TRD application.
- **Readout electronics** The standard readout for GEM detectors is typically based on an APV25 chip and measures the peak amplitude [? ]. A TRD needs additional information about the ionization along the track, to discriminate TR photons from the ionization of the charged particle. During the generic EIC R&D tests of GEM-TRD/T prototype we used a precise (125 MHz, 12 bit) FADC, developed at JLAB, with a VME-based readout. Those FADCs have a readout window (pipeline) of up to  $8 \mu\text{s}$ , which covers the entire drift time of the GEM-TRD/T prototype. The GAS-II pre-amplifier ASIC chips [? ] have been used, which provides  $2.6 \text{ mV/fC}$  amplification with a peaking time of 10 ns. A typical waveform signal, analyzed with the FADC

system is shown in Fig. ???. The flash ADC has a sampling rate of 125 MHz and 12 bit resolution but provides only pipe-lined triggered readout with a total price of about \$ 50 per channel. The collected high resolution data recorded in test beams allow us to estimate the minimum needed shaping times of preamplifier, the FADC sampling rate and corresponding resolution. Development of the new streaming readout, similar to described above, will be needed to enable the streaming of zero-suppressed data over fiber links. The currently employed readout electronics will be used to formulate a final set of specifications in driving the design of an ASIC and readout in conformance with the EIC streaming readout architecture.

- **TR-radiator**

A low mass radiator available for mass production is critical and various materials still need to be tested and optimized. This includes the optimization of a pseudo-regular radiator using thin ( $\sim 12 - 15 \mu\text{m}$ ) Kapton foils and thin net spacers and a detailed test of available fleece/foam materials for TR-yield.

- **Gas-system**

Over the past few years, the price of Xe has gone up significantly. Design and development of a recirculation system to purify, distribute, circulate, and recover the gas, possibly based on a design of ATLAS TRD gas system at CERN will be necessary, but will require only moderate R&D.

- **Data analysis and machine learning for  $e/\text{hadron}$  separation** To determine the electron identification efficiency and pion rejection power we tested several methods: total energy deposition, cluster counting, and a comparison of the ionization distribution along a path using maximum likelihood and neural network (NN) algorithms. The maximum likelihood and NN algorithms demonstrated similar performances. However, the NN algorithm has an advantage in practical application as it allows for the optimization of various test parameters and was used as the main analysis method. The ionization along the track was used as input to a neural network program (JETNET [? ], ROOT-based TMVA [? ]). Further development of Machine Learning algorithms for  $e/\pi$  separation is ongoing.

- **Machine learning on FPGA**

Our plan is to develop and build a functional demonstrator for FPGA Machine Learning application. A FPGA-based Neural Network application would offer real-time, low latency ( $1-4 \mu\text{s}$ ), particle identification. It would also allow for data reduction based on physical quantities during the early stages of data processing. This will allow us to control data traffic and offers the possibility of including detectors with PID information for online high-level trigger decisions, or online physics event reconstruction. The preliminary tests have been performed during the generic EIC R&D program.


### 5.3. *PICOSEC: MPGD for TOF at the EIC*

#### References

- [1] A. Braem et al., Nucl. Instr. and Meth. A 553 (2005) 187; H. Hoedlmoser et al., Nucl. Instr. and Meth. A 574 (2007) 28.
- [2] A. Breskin et al., Nucl. Instr. and Meth. A 367 (1995) 342; A. Di Mauro et al., Nucl. Instrum. Meth. A 371 (1996) 137; C. D. R. Azevedo et al., JINST 5 (2010)P01002; M.Alexeev et al., Nucl. Instr. and Meth. A 623 (2010) 129; J. Escada et al., J. Phys. D: Appl. Phys. 43 (2010) 065502.
- [3] The RD26 Collaboration, RD26 status reports: CERN/DRDC 93-36, CERN/DRDC 94-49, CERN/DRDC 96-20.
- [4] W. Anderson et al., Nucl. Instr. and Meth. A 646 (2011) 35.
- [5] J. Agarwala et al., Nucl. Instr. and Meth. A 936 (2019) 416.
- [6] M. Blatnik et al., IEEE Transactions on Nuclear Science 62 (2015) 3256.
- [7] B. Azmoun et al., IEEE Transactions on Nuclear Science 66 (2019) 1984.
- [8] A. Bondar et al., Nucl. Instr. and Meth. A 481 (2002) 200; F. Resnati et al., J.Phys.: Conf. Ser. 308 (2011) 012016.
- [9] L. Velardi, A. Valentini and G.Cicala, Diamond & Related Materials 76 (2017) 1.
- [10] J. Agarwala et al., Nucl. Instr. and Meth. A 952 (2020) 161967; C. Chatterjee et al., J. Phys.: Conf. Ser. 1498 (2020) 012008; F. M. Brunbauer et al., arXiv: 2006.02352.

### 6.1. Streaming readout at the EIC

The EIC readout architecture is shown in Fig. 3, where the readout chain is partitioned into three distinct functional profiles: FEB, FEP and DAQ.



The FEBs, Front-End Boards, are specifically designed for, mounted on or in close proximity to each of the sub-detectors. These boards conform to the geometry requirements of each of the sub-detectors and may be further constrained by power dissipation or heat loads, cooling services, radiation levels, cabling requirements and serviceability. The FEBs are characterized by the use of low noise, low power circuitry with analog front-ends and digitization, processing and drive capabilities. This high density mixed-mode circuitry is normally implemented with Application Specific Integrated Circuits (ASIC) when a very large number of readout channels justifies development efforts; alternatively, discrete implementations with commercial-off-the-shelf (COTS) components may be developed, if permitted by the established design boundaries.

The FEPs, Front-End Processors, are located outside of the proper sub-detector volume, or in close proximity, and interface to FEBs. The FEPs aggregate the data output streams from multiple FEBs and are designed to handle different types of FEBs, i.e., a few FEP designs with generic interfaces will handle multiple sub-detectors. Some sub-detectors



may not necessitate the use of FEPs, such as MAPS, as these may interface directly to switches, servers or link exchange modules (e.g., FELIX). The FEPs make extensive use of FPGAs for data processing, providing an opportunity to decrease the available output data bandwidth requirements by a factor of ten (10). This can be accomplished by various methods or algorithms, such as zero suppression or via ML and AI filtering. High performance FPGAs, available as COTS, provide the processing power, speed and flexibility of use. The choice of FPGAs and their location within the experimental area must be carefully considered with regards to radiation levels. As with the design considerations for the FEBs, the FEPs will play a critical role in configuration, control and timing of the various subsections of the readout chain via optical fibers.

The DAQ segment will consist of Front-End Link eXchange (FELIX) boards, servers and network switches, which will be located away from the detector, facilitated by the extensive use of optical fibers. Servers and switches are available as COTS. The FELIX board, originally developed for the ATLAS experiment, interfaces the DAQ and the detector front-end and functions as a router between custom serial links and a commodity switching network using standard technologies. FELIX is designed to be detector independent, supports the CERN GBT protocol to connect to front-end units, supports distribution of timing, trigger and control and supports calibration operations. It is expected that an updated design will be necessitated to fully benefit the EIC streaming readout. Optical fibers from the FEB and FEP are employed for data transport.

With the implementation of a fully streaming readout model, it is important to provide enough available bandwidth at the front-end (FEB) to handle potentially highly variable rates (signal/background/noise) from all detector systems. The total anticipated bandwidth limits for the whole detector are shown to be on the order of 100 Tbps at the front-end. With the implementation of the FEP stage we assume limits can be decreased by a factor of ten to 10 Tbps at the server/switch/link exchange level. Given current estimates on the physics interaction rates, following back-end processing and buffering, the expected data output rates to more permanent storage is anticipated to be around 100 Gbps. It is anticipated that the collision rate at the EIC will be four orders of magnitude lower than that at LHC: filtering out potential high background by means of ML/AI algorithms will be critical, however.

Timing, generally consists of global timing, busy, synchronization and clock distribution within the experimental area. Precision timing will require precision timing referencing the beam or beam crossings, as well as multiple control loops addressing stability and drift. A single accelerator clock reference may exhibit timing jitter of a few hundred femtoseconds, implemented with single-mode fibers and periodic phase correction.

Clock distribution jitter via multi-mode optical fibers, clock de-skewing and distribution via backplanes and employing COTS components can result in clock jitter of about 4 ps, which is sufficient for detectors requiring timing resolution in the 100 ps range. As some EIC timing detectors are specified for 20 ps timing resolution, it is expected that clock jitter will need to be better, within the 1 ps range.

Consideration for calibration, testing and timing of the various subsections and components of the EIC streaming readout requires that every part of the readout chain must be designed for the intended purpose of streaming readout and also for a triggered implementation. That is, the readout system should be able to operate with and without a trigger.

## *6.2. Streaming readout for MPGD detectors*

### *6.2.1. Online MPGD data reduction and filtering*

### *6.2.2. ML techniques for MPGD data processing*

## *6.3. MPGD front-end electronics adapted to triggerless mode*

### *6.3.1. Presently existing front-end ASICs*

Several front-end chips are presently existing to read MPGD detectors. Most of them are based on a charge-sensitive amplifier (CSA) associated to an amplifier/shaper. Signals are then either stored on a capacitor array or digitized and the data are stored in a digital buffer. The data or signals are then transferred to next element of the readout chain.

These chips are in general foreseen to be used in triggered mode, meaning that a trigger signal is delivered to the chip to trigger the readout of one event. This signal can be generated internally, based for instance on the amplified analog signals discrimination, or is transmitted from an external trigger system. In a streaming readout DAQ environment however no trigger signal is used and the front-end chip is meant to transmit to the DAQ system all signals which are read. Such a feature has to be foreseen in the chip architecture, and the output bandwidth of the chip has also to be large enough to transmit the foreseen data flux.

Only a few front-end ASICs compatible with a streaming readout DAQ are presently available. Among them the SAMPA chip and the VMM chip begin to be used in several particle physics experiment.

*SAMPA chip.* The SAMPA chip is a 32-channel ASIC with on-board pre-amplification, pulse shaping, digitization and DSP subsections, including a high-bandwidth digital interface for computer readout (11 links at 320 MB/s). It was developed by a group lead by the Sao Paulo University.

The chip is fabricated with TSMC 130 nm CMOS technology with a chip area of  $9.6 \times 9.0 \text{ mm}^2$ . A charge-sensitive amplifier amplifies the measured analog signals, followed by a near-Gaussian pulse shaper. The 10-bit Successive Approximation ADC digitizes the amplified and shaped signals at a sampling rate of 10 MS/s (which can be configured to up to 20 MS/s), whereas the on-board DSP circuitry filters and carries out signal processing and compression operations on the digitized data. The chip offers a sufficiently high gain of 20-30 mV/fC with a low-noise performance (less than  $1000 e^-$ ). The chip is well adapted to the readout of MPGD, in particular in the TPC (ALICE TPC at CERN) and in the front trackers.

*VMM chip.* The VMM chip was developed at Brookhaven (BNL) as a 64-channel mixed signal ASIC based on the IBM 130 nm technology for tracker readout, in particular the Micromegas and sTGC detectors of the ATLAS Muon Spectrometer's New Small Wheel upgrade. The chip features a novel third-order filter and shaper architecture, which results into a higher dynamic range, enabling the measurement to achieve a high resolution at moderate input capacitance ( $<200 \text{ pF}$ ), while it is also able to handle large capacitance up to 3 nF. The architecture offers a variable gain in eight values from 0.5 to 16 mV/fC with four possible shaping time intervals between 25 and 200 ns.

An excellent feature of the chip is to have both time and amplitude (peak) detection circuitry on-board. For each channel the amplified and shaped signal is passed over to both a peak detector and time detector working in tandem and giving their respective output to a digitization subsection. Output from the peak detector is given to both a 6-bit ADC for a dedicated low-delay output (50 ns delay), to be used for trigger or lower precision measurements, and to a 10-bit ADC for precision read-out, whereas the time detector has its output passed over to an 8-bit ADC for TDC functionality. Output data flux can reach a bandwidth of 1 Gbit/s.

### 6.3.2. Project for a new chip in 65 nm technology

A new initiative was recently launched by the University of Sao Paulo (Brazil) and the IRFU institute of the CEA Saclay (France) to develop a new front-end ASIC dedicated to almost all kinds of MPGD and beyond, and compatible with the requirements of modern streaming readout DAQ architectures, like the ones foreseen for the EIC project. The new ASIC, named SALSA chip, is meant to be versatile enough to cover most MPGD applications with different requirements, like large capacitance electrodes of large area track detectors, long time gates of time projection chambers, or low amplitude signals of photon detectors. Thus the ASIC shall be compatible with large ranges of signal amplitudes, electrode capacitances and propose a large range of peaking times, with optimized data processing. In order to propose such a versatile chip with a limited die size and a low power consumption, it will be designed in a more modern TSMC 65 nm technology, compared to the 130 nm technologies of the previous generation of chips.

*Preliminary specifications.* A preliminary set of specifications for SALSA chip was determined based on the requirements of the MPGD detectors foreseen in EIC project and in other experiments. The ASIC is foreseen to provide the readout of 64 channels adapted to a large range of electrode capacitance and with a large range of gains and

peaking times (Table 1), accepting both polarities of input signals. The front-end part of the channel integrates an optimized charge sensitive amplifier with tunable gain and anti-saturation circuit, a quasi-gaussian shaper with different selectable peaking times and an adapted pole zero cancellation stage. The output of the shaper is sampled and digitized thanks to an in-channel 50 MS/s 12 bits ADC.

Parameter	Value
Analog characteristics	
Number of channels	64
Peaking time range	50 to 500 ns
Input dynamic range	0-50 fC to 0-5 pC
Input capacitance range	Optimized for 200 pF, reasonable gain up to 1 nF
Input rates	25 kHz/channel, with faster CSA reset for larger rates
Additional feature	Reversible polarity
Digital characteristics	
ADC sampling rate	10 to 50 MS/s
ADC dynamics	12 bits
Data processing	Pedestal subtraction, common mode correction, zero suppression, peak finding, software trigger generation
Output data links	One or a few gigabit links

Table 1: Preliminary specifications of the SALSA chip

An integrated DSP will be able to process the data in order to perform basic treatments like pedestal subtraction, common mode noise correction and zero suppression. More sophisticated processing could be also done like trigger signal generation, peak finding, hit counting, or other kind of digital data filtering. Continuous readout mode compatible with a streaming DAQ will be naturally proposed, but the ASIC will be also compatible with more classic triggered DAQ. One or several gigabit output links will be available for the data transmission to the DAQ. The ASIC shall be able to stand input rates of at least 25 kHit/s/channel and beyond.

The SALSA chip die shall be small enough, around 1 cm<sup>2</sup>, to be able to be installed close to the detectors, with a low power consumption at the level of 15 mW/channel. Thanks to the use of 65 nm technology and hardening techniques (TMR) the SAMPA chip will be able to stand quite harsh environment in terms of particle radiation (TID and SEUs), magnetic field or temperature.

*Status and timeline.* The SALSA chip project was recently launched and the ASIC specifications will be soon fully finalized. Studies on the front-end and ADC architectures with the 65 nm technology are ongoing at Sao Paulo and Saclay. Some developments done at CERN on specific IP blocks for high-speed data links or internal PLL-generated clocks could be of interest in the ASIC architecture. Contacts are taken with CERN group about this topic. Development phase is foreseen to last around two to two and half years, followed by a preserial production phase of one year. Production will begin after this phase, in accordance with the EIC timeline.

### 6.3.3. Needs for specific MPGD applications

Some particular applications of gaseous detectors require readout electronics with specific characteristics which may be barely provided by readout ASICs described above. This is particularly true for the following MPGD applications.

*Photon detectors.* Gaseous photon detectors are foreseen to be used for instance in RICH detectors, in particular in EIC experiments. Cerenkov photons are converted in electrons in photosensitive material, like Cesium Iodide (CsI), deposited on specific glass windows. The low conversion yield, at the level of 10% induces very low signal amplitudes, close to the noise level. Readout electronics noise must be low enough in order to limit the degradation of the signal. Regular ASICs as listed above could have a low enough noise level compatible with the requirements of the photon detectors. But for more stringent specifications very low-noise amplification-only ASICs, like the IDeF-X chip [ref], could provide a solution, with noise levels at the order of 100 electrons. However these chips would have

limited functionalities and need to be associated to other chips which will take in charge the remaining of the signal treatment chain.

*ps-level time resolution detectors.* MPGD detectors can achieve excellent time resolutions, at the level of a few tens of ps, when they are associated with a Cerenkov radiator and a layer of photoemissive material, like in the PICOsec project [ref]. Present prototypes are using specific very fast discrete readout electronics, associated with ps-level time digitizers based for instance on the SAMPIC TDC chip [ref]. However these present electronics are not well integrated and require a quite large space to be installed. Projects to develop more integrated fast amplifiers and TDC are considered but this has not yet been achieved. An other solution would be to use present or in-development ASICs, but this could lead to a degradation of the time resolution of the detector as it would be limited by the sampling rates or the time resolution provided by these chips.