

# Status Setup CACTUS Milano-Bicocca

CACTUS@MILANO-BICOCCA

DUNE Photosensor WG Meeting (DUNE-SP-PDS)

October 17, 2023



- We have started the official test with delivery 2 box 9:
  - Today we finished tray#1 🎉
- So far the I-V measurements are good and in agreement with the 1st and 3rd thermal cycle.
- DCR values are low (18-39 mHz/mm<sup>2</sup>) .
- Thanks to Tommaso for helping us with the new version of Labview.
- In the next few days we hope to measure one tray per day.
- Should the I-V extended measurement be carried out for all SiPMs or only when we see a noisy SiPM?

