

Office of
Science

Contribution ID: 72

Type: **not specified**

Measurement of stray neutron radiation within a proton therapy facility

Tuesday, 29 April 2014 16:40 (25 minutes)

Proton therapy is an advanced radiation treatment that destroys cancer cells with high accuracy. Proton treatment beams can be fine-tuned to deposit most of their energy at a specific depth within the patient, keeping irradiation of adjacent organs and healthy tissue to a minimum. However, proton interactions with materials in the beam line and with patient's tissue create secondary neutrons with energies from thermal to several tens or hundreds of MeV. These secondary neutrons may increase the risk of a second cancer significantly, and should be taken seriously.

Dosimetry measurements around medical proton accelerators are very complex due to heterogeneous radiation fields, for this reason the use of well proved and well known measurement systems is necessary. In 2013, organised by EURADOS Working Group 9 (Radiation Protection Dosimetry in Medicine) a large set of intercomparison measurement was carried out in the IBA (230 MeV) active-scanning proton beam therapy facility in Trento, Italy. Several types of active detectors and dosimeters were used: Bonner Sphere Spectrometers (BSS), HAWK TEPC, several REM-counters such as WENDI-II, Berthold, RadEye, NM2B, and new HMGU electronic dosimeters. The proton beam was directed into a water phantom (30 x 30 x 60 cm³) with a 1 liter (10 x 10 x 10 cm³) target volume, simulating the patient. The isocenter of the beam was in depth of 15 cm inside the water phantom. Neutron ambient dose equivalents, $H(10)$, were measured at several positions inside the treatment room at four angles: 0°, 45°, 90°, and 135° with respect to the beam line direction, and at different distances.

A generally good agreement among $H(10)$ values within 30% was observed for the different types of used detectors. The highest neutron $H(10)$ value of about 60 $\mu\text{Sv/Gy}$ was measured along the beam axis at distance of 1.15 m from isocenter. However, the lowest neutron $H(10)$ value measured in the treatment room was about 1.1 $\mu\text{Sv/Gy}$ at distance of 4.25 m, and at 90° with respect to beam line.

Measurements with BSS provided valuable information on the spectral energy distribution of secondary neutrons and $H^*(10)$ at 4 different angles around the water phantom. Secondary neutron spectra generated by a proton beam impinging a water phantom were also calculated using Monte Carlo simulation with the GEANT4 code.

Dosimetry measurements in the phantom included different passive detectors (RPL, several types of TLDs and CR-39). Their future analysis will give insight in dose contributions from different radiation types depending on the position in the phantom and distance from the target volume.

The results of this large set of intercomparison measurements and Monte Carlo simulations may be valuable for medical physicists and radiation protection officers.

Primary author: MARES, Vladimir (Helmholtz Zentrum München, Institute of Radiation Protection, 85764 Neuherberg, Germany)

Co-authors: DOMINGO, Carles (Departament de Física, Universitat Autònoma de Barcelona, Bellaterra, Spain); AL-GRANATI, Carlo (Medical Physics Department, Trento Proton Therapy Center, Via Valentina Zambra, 16 - 38121 Trento, Italy); FROJD, Erik (CERN, The European Organization for Nuclear Research, CH-1211 Geneva 23, Switzerland); FELLIN, Francesco (Medical Physics Department, Trento Proton Therapy Center, Via Valentina Zambra, 16 - 38121 Trento, Italy); TROMPIER, François (Institute for Radiological Protection and Nuclear Safety, Human Health Division, BP17, 92260 Fontenay-aux-Roses; France); FARAH, Jad (Institute for Radiological Protection and Nuclear Safety, Human Health Division, BP17, 92260 Fontenay-aux-Roses; France); KUBANCAK, Jan (Department of Radiation Dosimetry, Nuclear Physics Institute, CZ-250 68 Řež, Czech Republic); SCHINNER, Konstantin (Helmholtz Zentrum München, Institute of Radiation Protection, 85764 Neuherberg, Germany); STOLARCZYK, Liliana (Institute of Nuclear Physics PAN, Radzikowskiego 152, 31-342 Krakow, Poland); KLODOWSKA, Magdalena

(Institute of Nuclear Physics PAN, Radzikowskiego 152, 31-342 Krakow, Poland); LISZKA, Malgorzata (Institute of Nuclear Physics PAN, Radzikowskiego 152, 31-342 Krakow, Poland); SCHWARZ, Marco (Medical Physics Department, Trento Proton Therapy Center, Via Valentina Zambra, 16 - 38121 Trento, Italy); WIELUNSKI, Marek (Helmholtz Zentrum München, Institute of Radiation Protection, 85764 Neuherberg, Germany); MAJER, Marija (Ruđer Bošković Institute, Bijenička c. 54, 10000 Zagreb, Croatia); ROMERO EXPÓSITO, María Teresa (Departament de Física, Universitat Autònoma de Barcelona, Bellaterra, Spain); PLOC, Ondrej (Department of Radiation Dosimetry, Nuclear Physics Institute, CZ-250 68 Řež, Czech Republic); OLKO, Pawel (Institute of Nuclear Physics PAN, Radzikowskiego 152, 31-342 Krakow, Poland); HARRISON, Roger (University of Newcastle upon Tyne, Newcastle, UK); MILJANIĆ, Saveta (Ruđer Bošković Institute, Bijenička c. 54, 10000 Zagreb, Croatia); TRINKL, Sebastian (Helmholtz Zentrum München, Institute of Radiation Protection, 85764 Neuherberg, Germany); GEORGE, Stuart Patrick (CERN, The European Organization for Nuclear Research, CH-1211 Geneva 23, Switzerland); DUFEEK, Vladimir (National Radiation Protection Institute, Bartoškova 28, 140 00 Praha, Czech Republic); RÜHM, Werner (Helmholtz Zentrum München, Institute of Radiation Protection, 85764 Neuherberg, Germany); KNEŽEVIĆ, Željka (Ruđer Bošković Institute, Bijenička c. 54, 10000 Zagreb, Croatia)

Presenter: MARES, Vladimir (Helmholtz Zentrum München, Institute of Radiation Protection, 85764 Neuherberg, Germany)

Session Classification: Session 4. Medical Accelerators, Convener: Vladimir Mares