



Contribution ID: 79

Type: **not specified**

Detection of engineered cerium oxide nanoparticles in soil

Wednesday, 11 March 2015 11:00 (20 minutes)

The detection of engineered CeO₂-NPs in complex natural media is very challenging due to the low expected CeO₂-NP concentrations and the comparatively high background of Ce-containing minerals of similar size range. We here present a new analytical method, based on single particle (sp) ICP-MS analysis for identification and quantification of engineered cerium oxide nanoparticles (CeO₂-NPs). We expect pulse signals of natural Ce-containing particles to be low and to not represent the true size of the particles detected by the sp-ICP-MS. In contrast, engineered CeO₂-NPs will appear as a spike which is significantly higher than the background signal and can be used to determine the mass and number concentration as well as the particle size of the CeO₂-NPs. Our hypothesis was tested with a set of experiments using CeO₂-NP-spiked natural colloid suspensions as well as colloidal extracts of a natural soil spiked with CeO₂-NPs. With our current single-isotope method in sp-ICP-MS we are able to detect the addition of the two-fold Ce-concentration compared to the background concentration in all types of samples. Multiple isotope techniques are currently investigated to extend our method by enabling the use of elemental ratios on single particle level and improving the detection limits considerably.

Primary author: VON DER KAMMER, Frank (University of Vienna)

Co-authors: PRAETORIUS, Antonia (University of Vienna); HOFMANN, Thilo (University of Vienna)

Presenter: VON DER KAMMER, Frank (University of Vienna)

Session Classification: 5B Tracking NM in complex matrices

Track Classification: Parallel session 5B: Tracking NM in complex matrices